CERTIFICATE

Of Conformity Low Voltage Directive 2014/35/EU

Registration No.: AT78251SC200078

Report No.: 78251SC20007801

Applicant	:	Wuhan Lotuxs Technology Co., Ltd. 5F, E2 Building, Future City, NO. 999 High-tech Avenue, East Lake High-tech Development Zone, Wuhan, 430206, China		
Product	:	Laser Therapy Device		
Identification		Model No.	:	PowerCure Pro, PowerCure
		Trade Mark	:	N.A.
		Rating	:	Input:5V/2A Output: 1300mW Battery: 3.7V, 5000mAh Internal operating frequency: 1Hz
Test Standards	:	EN 60335-1:2012+A11:2014+A13:2017 +A1:2019+A14:2019+A2:2019+A15:2021 EN 62233:2008		

The certificate of conformity is based on an evaluation of a sample of the above mentioned product. Technical report and documentation are at the applicant's disposal. This is to certify that the tested sample is in conformity with Low Voltage Directive 2014/35/EU relating to electrical equipment designed for use within certain voltage limits. The certificate does not imply assessment of the series-production of the product. The applicant of the certificate is authorized to use this certificate in connection with EU declaration of conformity specified in Article 15 and Annex IV of the Directive.

